


# VERIFICATION OF CALIBRATION REPORT

SEP 10 2009

Of DataMaster cdm BreathTest Instrument

State of Alaska

Scientific Crime Detection Laboratory -- Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130296</u>	
Name <u>John J. Waldron</u>		ID# <u>3392</u>	Date <u>08/14/09</u>
Agency <u>Unalaska Department of Public Safety</u>		Phone <u>907-581-1233</u>	
Instrument Location <u>Unalaska Department of Public Safety, PO Box 370, 29 Safety Way, Unalaska, AK 99685</u>			
Alco S/N <u>X172985</u>		Target Value <u>.083</u>	High Pressure <u>400</u>
Alco Test Values		<u>.083</u> 1 <sup>st</sup> Test Value	<u>.083</u> 2 <sup>nd</sup> Test Value
Signature <u>John J. Waldron</u>			
(OVER)			

BMB  
9/15/09


(Do not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 29<sup>th</sup> day of Sept, 2009.

  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



**VERIFICATION OF CALIBRATION REPORT**  
*Of DataMaster cdm BreathTest Instrument*  
*State of Alaska*  
*Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program*

DataMaster cdm S/N 130296

Supervisor/Operator & Number Performing the Verification Procedure: John J. Waldron, #3392  
Department and Date: Unalaska DPS,

(CONTINUED FROM FRONT PAGE)

**C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE  
DIAGNOSTIC CHECK IN THE MARKED BOXES.**

**NONDRINKING SUBJECT TEST**

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130296

AUGUST 14, 2009

OPERATOR'S NAME: WALDRON/JOHN/J  
OPERATOR'S NUMBER: 3392  
SUBJECT'S LAST NAME: VOC  
SUBJECT'S FIRST NAME/MI: VOC  
O.L. #: 01234567890  
DEPT/AGENCY: UNK1  
CASE/REPORT: 09-00000  
TEST TYPE: U  
ALCO TARGET VALUE: .083  
ALCO S/N: 1172955

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.80 in		
ALCO TARGET	.082	17:03
BLANK TEST	.000	17:04
INTERNAL STANDARD	VERIFIED	17:04
ALCO TV 29.78 in	.083	17:04
BLANK TEST	.000	17:05
SUBJECT SAMPLE	.000	17:05
BLANK TEST	.000	17:06
ALCO TV 29.80 in	.083	17:07
BLANK TEST	.000	17:07

**DIAGNOSTIC CHECK**

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130296

AUGUST 14, 2009  
TIME 17:08

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
SAPOMETER:	29.80 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

"#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKL/MNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~■